

Modificationcard for: EG 750 f.4Press. Celis control**Part No. : 316820****Print Date: 07-11-02**

| DATE | Lev. | + | Tekst (DK). | Description |
|-------------------|-------------|----------|--|---|
| 15-11-00 | 1 | | 1. maskine som 316800 niv 8 Forsynet med MK3 elektronik CPU 305116 Panel 305316 Software 2,0 eller højere | 1 unit as 316800 mod. Level 8 MK3 electronics CPU 305116 Panel 305316 Software 2,0 or higher |
| 7/12-00 | 2 | + | R92 forbundet til Q14 kortsluttet R129 forbundet til Q15 kortsluttet For at få mere moment og speed på conv. motor (bruges sammen med IRF610 transistor | R92 connected to Q14 short circuited R129 connected to Q15 short circuited To obtain more conveyor motor speed and torque |
| 29/3-01 | 3 | + | Software 2.07 med Dev , Fix , dry heat prioritet, 1KW tilsluttet i upper og 2KW i lower dry | Software 2.07 with Dev , Fix , dry heat prioritet. 1KW connected in upper and 2KW in lower dry |
| 17/4-01 5/1-01 | 4 | + | 3uF kondensator nr.29005 i serie med udsugningsblæser for at sænke kapacitet og nedsætte fordampning. Fra 5/1-01 Drivaxel dry rack flyttet 24mm mod vandrack, sidebox 316181 modificeret se tekn info # 30 | 3uF capacitor part no. 29005 in series with exhaust blower to reduce capacity and reduce evaporation from tanks. Since 5/1-01 Driveaxle dry rack moved 24mm towards washrack, sidebox 316181 modified to give room for axle. |
| 18/6-01 | 5 | | Blæsekanal 316698 med 3x1000W varmelegeme 305148. 1, 2 eller 3Kw forbundet iflg diagram | Blower channel 316698 with heating element 3x1000W 305148, 1, 2 or 3Kw connected according to diagram |
| 20/8-01 | 6 | + | Dev rack 75cm R2 w guide 316487 Wash/Fix rack 75cm R2 w guide 316498 Forbedret transport vdr. Ridser på bløde film | Dev rack 75cm R2 w guide 316487 Wash/Fix rack 75cm R2 w guide 316498 Improved transportation ref. scratches on soft film base |
| 4/10-01 | 7 | + | En modstand på 68 Kohm monteres parrallelt over R108 som er på 10 Kohm. Senere bliver R108 ændret til 8,87 Kohm For at undgå at drivmotoren stopper hvis forsynings spændingen er høj (f.eks 237V og trafo står på 208v). tekn. info #34 | A resistor 68Kohm connected parrallel to R108 which is 10Kohm. Later R108 is changed to 8.87Kohm. Problem were drive motor stopping occasionally at high supply voltage. (E.g. input 237V. and trafo connected for 208V) |
| 24/10-01 | 8 | + | Kabelsæt til dry relæer 305349 skal monteres ekstra flex 809110 ved spadsko. Trafo, to hvide ledninger forsynes med flex 809101. | Cableset for dryer relays 305349 fitted with extra flex 890110 at spade connectors. Trafo, the 2 white wires fitted with flex 809101. |
| 5/01-02 | 9 | | Drain system med silicone slanger uden spændebånd | Drain system with silicone hoses, no hose clamps fitted |
| 25/3-02 | 10 | . | Iwaki circulation pumps afløser spal 68215 | Iwaki circulation pumps |
| 1/5-02 | 11 | | 3-Fase relæ print 305219/305296, erstattet med 1-fase relæ print 305297 + 13 konnektorer for tørrer sektioner. 8489 + 8490 Interconnection diagram 6355 + 6359 | 3-Phase relay PCB. 305219/305296 replaced with 1- phase relay PCB 305297 + 13 connector for dryer sections. 8489 + 8490 Interconnection diagram 6355 + 6359 |
| 12/08-02 | 12 | + | Nyt CPU PCB MK4 305216 med feedback på conv. motor styring | New CPU PCB MK4 305216 improved conveyor motor control with feedback |
| 12/08-02 | 13 | | Ny 3 fase main sw box 305548 | New 3 phase main sw. box 305548 |
| 16/08-02 | 14 | | Dry rack med blå 78T tandhjul for bedre transport af lange film | Dryer rack with blue 78T gearwheel to optimize transportation of long films |

